

# A4XP Beads Quality Control Report

## QC Item A: Macro-Scale Mixed DNA Marker Size Selection and Fragment Recovery

This test evaluates DNA fragment recovery and exclusion behavior of A4XP beads under different bead-to-sample ratios using mixed DNA markers.

Parameter	Description
Sample	185 $\mu$ L purified water + 5 $\mu$ L DL2000 DNA Marker + 10 $\mu$ L 50 bp DNA Marker
Test ratios	0.6x, 0.8x, 1.0x, 1.2x, 1.4x, 1.6x and 1.8x bead volume
Evaluation method	Agarose gel electrophoresis

Acceptance criteria:

- 0.6x: partially recover the 500 bp fragment; recovery of 750–2000 bp fragments should exceed 60%; QC and control batch band patterns should be consistent.
- 0.8x: partially recover the 250 bp fragment; recovery of 400–2000 bp fragments should exceed 60%; QC and control batch band patterns should be consistent.
- 1.0x: recover the 200 bp fragment and most of the 250 bp fragment; QC and control batch band patterns should be consistent.
- 1.2x: partially recover the 150 bp fragment; QC and control batch band patterns should be consistent.
- 1.4x: recover most of the 150 bp fragment, with no or very little 100 bp recovery; QC and control batch band patterns should be consistent.
- 1.6x: recover a small portion of the 100 bp fragment; recovery of fragments 150 bp and above should exceed 60% and approach or exceed 80%; QC and control batch band patterns should be consistent.
- 1.8x: recover the 100 bp fragment with recovery approaching or exceeding 80%; QC and control batch band patterns should be consistent.




Figure 1. Macro-scale mixed DNA marker size selection using A4XP beads from 0.6x to 1.8x bead ratios.

**Conclusion:** No obvious difference was observed between the QC batch and the control batch. The QC batch passed this test.

### QC Item B: Microscale Mixed DNA Marker Recovery at Different Bead Ratios

This test evaluates microscale recovery of mixed DNA markers at selected bead ratios. The sample consisted of 197  $\mu\text{L}$  purified water, 1.5  $\mu\text{L}$  DL2000 DNA Marker and 1.5  $\mu\text{L}$  50 bp DNA Marker. The acceptance criterion was that the recovery difference between the QC batch and the control batch should not exceed 3% under the tested ratios.

Batch	Ratio	Qubit (ng/ $\mu\text{L}$ )	Yield (ng)	Recovery	Average	Difference vs. Control	QC Result
Original sample	-	1.64	164	-	-	-	-
Control	0.6x	0.570 / 0.588	57.0 / 58.8	34.76% / 35.85%	35.3%	-	-
	0.9x	0.936 / 0.932	93.6 / 93.2	57.07% / 56.83%	57.0%	-	-
	1.2x	1.12 / 1.12	112 / 112	68.29% / 68.29%	68.3%	-	-
	1.8x	1.44 / 1.41	144 / 141	87.80% / 85.98%	86.9%	-	-
QC	0.6x	0.504 / 0.488	50.4 / 48.8	30.73% / 29.76%	30.2%	-5.1%	Not passed*
	0.9x	0.984 / 0.930	98.4 / 93.0	60.00% / 56.71%	58.4%	1.4%	Pass
	1.2x	1.18 / 1.14	118 / 114	71.95% / 69.51%	70.7%	2.4%	Pass
	1.8x	1.43 / 1.42	143 / 142	87.20% / 86.59%	86.9%	0.0%	Pass

**\*Note:** At 0.6x, the QC batch recovery was lower than the control batch. The report notes that this may have been affected by small elution-volume variation under the 50  $\mu\text{L}$  elution condition. Future QC testing may use 100  $\mu\text{L}$  elution volume to reduce volume-related error.

**Conclusion:** The QC batch showed no obvious difference from the control batch overall and passed QC evaluation, with the 0.6x result noted for elution-volume-related variation.

### QC Item C: Background Nucleic Acid Assessment

This test evaluates whether the A4XP beads contain detectable background nucleic acid.

Batch	Replicate	Qubit Reading (ng/ $\mu\text{L}$ )	QC Result
QC batch	1	Too low	Pass
QC batch	2	Too low	Pass
QC batch	3	Too low	Pass
QC batch	4	Too low	Pass
Control batch	1	Too low	-
Control batch	2	Too low	-
Control batch	3	Too low	-
Control batch	4	Too low	-

**Conclusion:** Qubit readings were below the detectable range, indicating no detectable background nucleic acid under the tested conditions.